



LINEAR REGRESSION EQUATION

Zn ICP-MS, in µg/L = .538 + .995 * Zn clCP-MS, in µg/L; $R^2 = .999$

Confidence Intervals
Zn ICP-MS, in µg/L vs. Zn clCP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	.538	-.560	1.637
Zn clCP-MS, in µg/L	.995	.983	1.008